## Ficha Técnica TERMOMETROS

Producto	Termometro Infrarrojo
Unidades Temperatura	2 - 5.9 cms
Rango Temperatura	Temperatura Corporal: 32°C - 42.5°C
	Temperatura Objetos: 0°C - 100°C
Temperatura Ambiente	10°C - 40°C
Almacenamiento	0°C - 50°C
Pilas	2 x AA
Apagado Automatico	7 segundos
Precision	99,80%
Tolerancia	cercano a 0,3%

\* no incluye pilas



## **Producto y Packaging TERMOMETROS**





## **Certificados TERMOMETROS**





## CERTIFICATE OF CONFORMITY

No.: PTC20031004501S-LD01

Zhangve Nuoan Medical Equipment Co. Ltd Applicant

Address No.2, Intelligent terminal industrial zone A, Binghe new district,

Ganzhou district, Zhangye City, Gansu province Manufacturer Zhangye Nuoan Medical Equipment Co.,Ltd

Address No.2, Intelligent terminal industrial zone A. Binghe new district,

Ganzhou district, Zhangye City, Gansu province Non contact infrared thermometer

Product Trade Mark N/A

Model T-01

Power Rating Battery: DC 3.0V

The submitted sample of the above equipment has been tested and found to comply with the following European Directive:

Low Voltage Directive - 2014/35/EU

The standard(s) used for showing compliance with the essential requirements:

Applicable Standard(s)	Test Report(s) Number
EN 60601-1: 2006 + A11:2011 + A1:2013 + A12:2014	PTC20031004501S-LD01

This certificate is part of the full test report(s) and should be read in conjunction with it. This certificate is based on an evaluation of one sample of above mentioned product. It does not imply assessment of the production of the product. Without the written approval of Dongguan Precise Testing Service Co., Ltd, this certificate is not permitted to be reproduced, except in full. It is not permitted to use the test lab's logo. The CE marking may only be used if all the relevant and effective European Directives are applicable





Jacky Ou Manager Date: March 27, 2020





